

BULLARD COVX Quick Start Guide

Turn On

Press & Hold

**POWER
MODE**

Past Due Bump/CAL

To Bump / CAL

**POWER
MODE**

Confirm & Use



AIR

CAL 
BP - LMT 

CAL 
CAL - LMT 

**Normal
Operation**

After brief
warm-up period

CO 
0 ppm 
10:42

Fresh Air

Press & Hold

AIR

Turn Off

Press & Hold

**POWER
MODE**

⚠ WARNING

FOLLOW ALL INSTRUCTIONS AND WARNINGS IN THE COVX CO MONITOR USER MANUAL. FAILURE TO FOLLOW ALL INSTRUCTIONS AND WARNINGS CAN RESULT IN DEATH OR SERIOUS INJURY.